

## Data Sheet

<b>Type of sensor</b>	Planar Oxygen-Sensitive Spot
<b>Ordering code</b>	SP-PSt5-NAU-D5-YOP_US
<b>Article number</b>	200001096
<b>Batch number</b>	<b>SP-PSt5-1547-03</b>
<b>Production Date</b>	07 Dec 2015
<b>Reference devices</b>	SEBX 000 4 000 200

**Dear customer,**

Please upload the file with the BATCH NUMBER above (corresponding to the batch number printed on the package of your plate) in the SDR software. The file will be sent to you by e-mail. Choose this batch number in the software for your measurement.

**Calibration Results**

Atmospheric pressure:  hPa

	Phase signal	Temperature
	[°]	[C°]
<b>cal 0</b>		
<b>0% air sat</b>	56,14	10,3
<b>cal 100</b>		
<b>100 % air sat</b>	50,28	10,3
<b>cal 0</b>		
<b>0% air sat</b>	54,45	25,3
<b>cal 100</b>		
<b>100 % air sat</b>	47,05	25,3

Sensors are visually inspected.